




**BIOLOX®*delta***  
Ceramics in Orthopaedics

## **BIOLOX®*delta*** CERAMIC AT A GLANCE



- Excellent biological behaviour
- Low immunological response
- High wear resistance
- Reduced bacterial adhesion
- Safe in terms of metal ion release
- Best clinical outcomes for ceramic bearings
- High potential for cost-effectiveness of ceramic bearings in THA





**BIOLOX®*delta***

Ceramics in Orthopaedics

**BIOLOX®**  
MATERIAL  
MATTERS®

## References

- Faye PA, Roualdes O, Rossignol F, Hartmann DJ, Desmoulière A. Engulfment of ceramic particles by fibroblasts does not alter cell behavior. *Biomed Mater.* 2017;12(1):015023. doi: 10.1088/1748-605X/aa5aa2.
- Cunningham BW, Hallab NJ, Hu N, McAfee PC. Epidural application of spinal instrumentation particulate wear debris: a comprehensive evaluation of neurotoxicity using an in vivo animal model. *J. Neurosurg. Spine.* 2013;19(3):336-350. doi:10.3171/2013.5.SPINE13166.
- Savarino L, Baldini N, Ciapetti G, Pellacani A, Giunti A. Is wear debris responsible for failure in alumina-on-alumina implants? *Acta Orthop.* 2009;80(2):162-167. doi:10.3109/17453670902876730.
- Asif IM, Williams S, Fisher J, Al-Hajjar M, Anderson J, Tipper JL. *Front. Bioeng. Biotechnol.* 00793. doi:10.3389/conf.FBIOE.2016.01.00793.
- Asif I M. *Characterisation and Biological Impact of Wear Particles from Composite Ceramic Hip Replacements.* [PhD thesis]. Leeds, UK: University of Leeds; 2018. <http://etheses.whiterose.ac.uk/20563/>.
- Krenn V, Thomas P, Thomsen M, et al. Histopathological Particle Identification (The Krenn Particle Algorithm). *CeraNews* 2013;2:12-17.
- Grupp TM, Holderied M, Mulliez MA, et al. Biotribology of a vitamin E-stabilized polyethylene for hip arthroplasty – Influence of artificial ageing and third-body particles on wear. *Acta Biomater.* 2014;10:3068-3078. doi:10.1016/j.actbio.2014.02.052.
- Nikolaou VS, Edwards MR, Bogoch E, Schemitsch EH, Waddell JP. A prospective randomized controlled trial comparing three alternative bearing surfaces in primary total hip replacement. *J Bone Joint Surg Br.* 2012;94-B(4):459-465. doi:10.1302/0301-620X.94B4.27735.
- Piconi C, Porporati AA, Streicher RM. Ceramics in THR Bearings: Behavior under Off-Normal Conditions. *Key Engineering Materials.* 2014;631:3-7. doi:10.4028/www.scientific.net/kem.631.3.
- Trampuz A, Maiolo EM, Winkler T, Perka C. Biofilm formation on ceramic, metal and polyethylene bearing components from hip joint replacement systems. *Orthopaedic Proceedings.* 2016;98-B(SUPP 10):80-80. doi:10.1302/1358-992X.98BSUPP\_10.ISTA2015-080.
- Sorrentino R, Cochis A, Azzimonti B, et al. Reduced bacterial adhesion on ceramics used for arthroplasty applications. *J Eur Ceram Soc.* 2018;38(3):963-970. doi:10.1016/j.jeurceramsoc.2017.10.008.
- Beraudi A, Stea S, De Pasquale D, et al. Metal ion release: also a concern for ceramic-on-ceramic couplings? *Hip Int.* 2014;24(4):321-326. doi:10.5301/hipint.5000132.
- Kretzer JP, Mueller U, Streit MR, et al. Ion release in ceramic bearings for total hip replacement: Results from an in vitro and an in vivo study. *Int Orthop.* 2018;42(1):65-70. doi:10.1007/s00264-017-3568-1.
- Kocagoz SB, Underwood RJ, MacDonald DW, Gilbert JL, Kurtz SM. Ceramic heads decrease metal release caused by head-taper fretting and corrosion. *Clin Orthop Relat Res.* 2016;474(4):985-994. doi:10.1007/s11999-015-4683-1.
- Thomas P, Stea S. *Metal Implant Allergy and Immuno-Allergological Compatibility Aspects of Ceramic Materials.* Heidelberg, Germany: Springer-Verlag Berlin Heidelberg; 2015.
- Sharplin P, Wyatt MC, Rothwell A, Frampton C, Hooper G. Which is the best bearing surface for primary total hip replacement? A New Zealand Joint Registry study. *Hip Int.* 2017;28(4):352-362. doi:10.5301/hipint.5000585.
- Peters RM, Van Steenbergen LN, Stevens M, Rijk PC, Bulstra SK, Zijlstra WP. The effect of bearing type on the outcome of total hip arthroplasty. *Acta Orthop.* 2018;89(2):163-169. doi:10.1080/17453674.2017.1405669.
- Higuchi Y, Seki T, Hasegawa Y, Takegami Y, Morita D, Ishiguro N. 32-mm ceramic-on-ceramic total hip arthroplasty versus 28-mm ceramic bearings: 5- to 15-year follow-up study. *Hip Int.* 2019;29(1):65-71. doi:10.1177/1120700018760971.
- Toni A, Giardina F, Guerra G, et al. 3rd generation alumina-on-alumina in modular hip prosthesis: 13 to 18 years follow-up results. *Hip Int.* 2017;27(1):8-13. doi:10.5301/hipint.5000429.
- Kim YH, Park JW, Kulkarni SS, Kim YH. A randomised prospective evaluation of ceramic-on-ceramic and ceramic-on-highly cross-linked polyethylene bearings in the same patients with primary cementless total hip arthroplasty. *Int Orthop.* 2013;37(11):2131-2137. doi:10.1007/s00264-013-2036-9.
- Wyles CC, McArthur BA, Wagner ER, Houdek MT, Jimenez-Almonte JH, Trousdale RT. Ceramic femoral heads for all patients? An argument for cost containment in hip surgery. *Am J Orthop (Belle Mead NJ).* 2016;45(6):E362-E366.
- Carnes KJ, Odum SM, Troyer JL, Fehring TK. Cost analysis of ceramic heads in primary total hip arthroplasty. *J Bone Joint Surg Am.* 2016;98(21):1794-1800. doi:10.2106/JBJS.15.00831.
- Kurtz SM, Lau E, Baykal D, Odum S, Springer BD, Fehring TK. Are ceramic bearings becoming cost-effective for all patients? *J Arthroplasty.* 2018;33(5):1352-1258. doi:10.1016/j.arth.2017.12.011.
- Kurtz SM, Lau E, Baykal D, Odum SM, Springer BD, Fehring TK. Are ceramic bearings becoming cost-effective for all patients within a 90-day bundled payment period? *J Arthroplasty.* 2019;34(6):1082-1088. doi:10.1016/j.arth.2019.01.074.
- Product pictures: CeramTec GmbH, 2020.

